

AMENDMENT NO. 3 APRIL 2023
TO
IS 5158 : 1987 SPECIFICATION FOR PHTHALIC ANHYDRIDE , TECHNICAL
(Second Revision)

(Page 3, Appendix A, clause A-2.1.3.1) — Substitute following for the existing clause:

‘A-2.1.3.1 The thermometer shall bear a valid calibration certificate from any institution authorised to issue calibration certificate traceable to international or national measurement standards.’

Price Group 1